

UNCLASSIFIED//

RTTUZYUW RUOIAAZ0142 0941933-UUUU--RHMCSUU.

ZNR UUUUU

R 051140Z APR 13 ZYB

FM COMNAVSAFECEN NORFOLK VA

TO ALSAFE

BT

UNCLAS

ALSAFE 021/13

SECINFO/U/-//

MSGID/GENADMIN/COMNAVSAFECEN/20/MAR//

SUBJ/FLUKE PRODUCT SAFETY ALERT AND RECALL// POC/PERFETTO,

JOSEPH/CIV/UNIT:COMNAVSAFECEN/NAME:NORFOLK

/TEL:757-444-3520 X7264/TEL:DSN 564-3520 X7264 /EMAIL:JOSEPH.PERFETTO@NAVY.MIL//

GENTEXT/REMARKS/1. IN COOPERATION WITH THE CONSUMER PRODUCTS SAFETY COMMISSION, FLUKE HAS ISSUED RECALLS FOR FIVE OF ITS PRODUCTS.

2. PER THE FLUKE ANNOUNCEMENT CONSUMERS SHOULD IMMEDIATELY STOP USING THE DEVICES LISTED AND GO TO [HTTP://WWW.FLUKE.COM/FLUKE/USEN/](http://www.fluke.com/flukenet/support/safety/) SUPPORT/SAFETY/ FOR DETAILS ABOUT THE ITEMS INVOLVED AND THE SPECIFIC FOLLOW-UP ACTIONS FOR EACH.

3. THE PRODUCTS (FIVE TOTAL):

A. FLUKE 373, 374, 375, AND 376 DIGITAL CLAMP METERS B. FLUKE 190 SCOPEMETER C. FLUKE T2 ELECTRICAL TESTER D. FLUKE 333, 334, 335, 336, AND 337 DIGITAL CLAMP METERS E. FLUKE 1AC-1 VOLTALERT VOLTAGE TESTER 4. HAZARD IS EXPOSURE TO ELECTRICAL SHOCK, ARC FLASH AND/OR ELECTROCUTION DEPENDING UPON THE EQUIPMENT AND VOLTAGE INVOLVED. THE VOLTAGE DISPLAYED BY EACH OF THESE DEVICES MAY BE LESS THAN THE VOLTAGE IN THE ITEM BEING EVALUATED. IN SOME CASES, ZERO VOLTS WILL BE DISPLAYED WHEN A VOLTAGE IS PRESENT.

5. ACTION: PERSONNEL USING THESE FLUKE PRODUCTS SHOULD STOP USING THEM IMMEDIATELY AND TAKE FOLLOW-UP ACTIONS AS PRESCRIBED BY FLUKE.

6. ADDITIONAL INFORMATION IS AVAILABLE FROM THE FLUKE WEBSITE:

[HTTP://WWW.FLUKE.COM/FLUKE/USEN/SUPPORT/SAFETY//](http://www.fluke.com/flukenet/support/safety/)

BT

#0142